

FRX AED MAINTENANCE CHECKLIST

Date								
AED Unit Clean, no dirt or contamination; no damage present								
Supplies Available: a. 2 sets defibrillation pads, within expiration date, undamaged b. 1 Pediatric Key c. Ancillary supplies: <ul style="list-style-type: none"> • Towel • Scissors • Razor • CPR mask • Gloves d. Store unopened battery within install before date								
Status Indicator:								
Inspected by:								
Remarks, Problems, Corrective Actions:								

Prepared by Emergency Skills, Inc.
 Daily, check AED status indicator.
 Monthly, perform thorough check, Date and initial upon completion.
 Please make copies of this form as needed. Store completed forms in a safe place.
 To order supplies, call the Dept of Education - Sherry Fusco (718-391-8382), Celeste McGee (718-391-8566) or
 Donovan Lauther at (718) 391-8227.